

# SAFIRA™

Giving Anesthesiologists Control

## KEY POINTS TO REMEMBER

- Remove the black cap off the driver before use and dispose
- Remove the protective cap off the syringe
- Never store the driver with a syringe attached